

Showerbuddy Global Ltd.

PO Box 494, Waikanae, 5250 New Zealand

25th February 2013

Dear Barry,

Reference: review of the Showerbuddy Global Limited SB2 TubBuddy mobile shower chair, our Laboratory reference is 491835.

This letter is to advise you that this Laboratory has finished reviewing your product to the following Standard method: AS/NZS 3973:2009 *Shower / toilet chairs, Mobile and Static* (except materials and surface finishes).

The official nine (9) page report, dated 25th February 2013 and referencing job number 491835, identifies the sample, the test documents and the testing facility. All conformances with the AS/NZS standards are indicated in the report.

The Showerbuddy Global Limited SB2 ShowerBuddy Mobile shower chair, as represented by the test sample, including the supplied documentation and labelling as stated in the test report, complies with the relevant requirements of AS/NZS 3973:2009 *Shower / toilet chairs, Mobile and Static* (excluding the methods indicated in the report as "not tested") for user mass 160 kg.

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel Senior Test Technician, Test Laboratory Authorised signatory



